



Refrigerator Temperature Log for Vaccines



Strengthening the Vaccine Provider Network

Month: _____

Days: _____ to _____

Year: _____

Place an X in the box that corresponds with the temperature for morning and afternoon. Enter initials in the corresponding area. **Shaded zones represent unacceptable temperature ranges.**



If the recorded temperature falls outside the acceptable temperature range:

1. Put a "Do not use vaccine" sign on the affected vaccines.
2. Store the affected vaccines under appropriate conditions as soon as possible.
3. Notify your vaccine coordinator, or call the immunization program at your state or local health department for guidance.
4. Record the actions you take.

	SUNDAY		MONDAY		TUESDAY		WEDNESDAY		THURSDAY		FRIDAY		SATURDAY	
TIME	AM	PM	AM	PM	AM	PM	AM	PM	AM	PM	AM	PM	AM	PM
≥ 49°F														
48°F														
47°F														
46°F														
45°F														
44°F														
43°F														
42°F														
41°F														
40°F														
39°F														
38°F														
37°F														
36°F														
35°F														
34°F														
33°F														
≤ 32°F														
MIN/MAX Temp*														
Room Temp														
Staff Initials														

*Record MIN and MAX temperatures from the thermometer before resetting.
Check the Prescribing Information on the vaccine manufacturer's website for specific storage temperature instructions.

Adapted from the Immunization Action Coalition Temperature log for refrigerator. www.immunize.org/catg.d/p3037F.pdf. Accessed May 10, 2018.



Copyright © 2019 Merck Sharp & Dohme Corp., a subsidiary of **Merck & Co., Inc.** All rights reserved.

US-NON-03155 11/19



Refrigerator Temperature Log for Vaccines



Strengthening the Vaccine Provider Network

Month: _____

Days: _____ to _____

Year: _____

Place an X in the box that corresponds with the temperature for morning and afternoon. Enter initials in the corresponding area. **Shaded zones represent unacceptable temperature ranges.**



If the recorded temperature falls outside the acceptable temperature range:

1. Put a "Do not use vaccine" sign on the affected vaccines.
2. Store the affected vaccines under appropriate conditions as soon as possible.
3. Notify your vaccine coordinator, or call the immunization program at your state or local health department for guidance.
4. Record the actions you take.

	SUNDAY		MONDAY		TUESDAY		WEDNESDAY		THURSDAY		FRIDAY		SATURDAY	
TIME	AM	PM	AM	PM	AM	PM	AM	PM	AM	PM	AM	PM	AM	PM
$\geq 11^{\circ}\text{C}$														
10°C														
9°C														
8°C														
7°C														
6°C														
5°C														
4°C														
3°C														
2°C														
1°C														
0°C														
$\leq -1^{\circ}\text{C}$														
MIN/MAX Temp*														
Room Temp														
Staff Initials														

*Record MIN and MAX temperatures from the thermometer before resetting.
Check the Prescribing Information on the vaccine manufacturer's website for specific storage temperature instructions.

Adapted from the Immunization Action Coalition Temperature log for refrigerator. www.immunize.org/catg.d/p3037F.pdf. Accessed May 10, 2018.